

Portable Indoor Air Quality Instruments and Smart Sensor Specifications



Critical Environment Technologies Canada Inc. Unit 145, 7391 Vantage Way Delta, BC V4G 1M3 Canada IF: +1.877.940.8741
I: +1.604.940.8741
F: +1.604.940.8745
© April 2015 | Data subject to change without notice.

www.critical-environment.com

شرىت **آلاپرداز مميط** تهران - خیابان آزادی - مابین خیابان شادمهر و بزرگراه یادگار امام - شماره ۴۱۷ - واحد ۵ تلفن: E-mail: info@apm-co.ir ۰۲۱-۶۶۰۲۸۱۷۲-۵ فکس: Website: www.apm-co.ir ۰۲۱-۶۶۰۲۰۵۰۹

Table of Contents

Product Overview: Indoor Air Quality Instruments	4
Overview: Gas Detection Sensor Technologies	5
Target Gas Sensor Specifications	
RelativeHumidity(RH)	7
Temperature	7
Ammonia (NH ₃)	8
Arsine (AsH,)	8
Carbon Dioxide (CO ₃)	9
Carbon Monoxide (CO)	
Chlorine (Cl,)	13
Chlorine Dioxide (Cl0,)	
Combustibles	15
Ethylene (C,H,)	16
Ethylene Oxide (C,H ₄ O)	
Fluorine (F,)	
Formaldehyde (CH,0)	
Hydrogen(H ₁)	19
Hydrogen Chloride (HCI)	
Hydrogen Cyanide (HCN)	20
Hydrogen Fluoride (HF)	21
Hydrogen Sulphide (H,S)	22
Nitrogen Dioxide (N0,)	22
Nitric Oxide (NO)	
Oxygen (0,)	24
Ozone (0,)	25
Phosphine (PH3)	26
Silane (SiH ₄)	
Sulphur Dioxide (SO ₂)	28
Total Volatile Organic Compounds (TVOC)	
AdditionalChemicallSymbols	30
Important Notes	30

PRODUCT OVERVIEW Indoor Air Quality Instruments

YESAIR Eight Channel Air Quality Monitor



- » 8 sensor capacity
- » > 30 plug & play sensor choices
- » Data logging to SD flash card
- » Lightweight, contoured & comfortable handheld device



YES Plus LGA Fifteen Channel Air Quality Monitor



- » 15 sensor capacity
- » > 30 plug & play sensor choices
- » Data logging to SD flash card
- » Internal sample pump with inline filter





OVERVIEW Gas Detection Sensor Technologies

Categories of gas detection systems are defined by the technology they use: electrochemical sensors generally detect toxic gases, catalytic and infrared sensors detect combustible gases and TVOC sensors detect toxic gas, organic compounds and chemicals.



ELECTROCHEMICAL

Toxic gas sensors & oxygen sensors

LIFE SPAN

2 - 5 years (sensor type & manufacturer dependent). Oxygen typically 3 years.

SPECIFICITY

Specific to target gas with known cross sensitivity to a small variety of gases.

RANGE

Typically 0 - 1.0 ppm or 0 - 2,000 ppm, sensor dependent.

POISONING / DAMAGING

- Oxygen depravation
- Exposure to high concentrations of solvent vapours
- Very high concentrations of target gas
- Reactive gases
- Environments with high temperatures, low temperatures (freezing)
- Very low levels of relative humidity (less than 10 - 15%)

APPLYING SPAN GAS

- Use span gas with air balance or nitrogen balance
- Flow rate should be a minimum of 0.5 LPM (lighter-than air gases) to a maximum of 1.0 LPM (heavier-than-air gases)
- Do NOT humidify span gas when flowing

CATALYTIC

Combustible gas sensors, toxic gas sensors at very high concentrations (% volume)

LIFE SPAN

3 - 8 years (typically if not poisoned)

SPECIFICITY

Specific to combustible gases only in the LEL ranges

RANGE

0 - 100% LEL of target gas

POISONING / DAMAGING

High concentrations of target gas, lead vapours, silicon vapours, alkylated heavy metals.

APPLYING SPAN GAS

- Use span gas with air balance ONLY.
- Flow rate should be a minimum of 0.5 LPM (lighter-than air gases) to a maximum of 1.0 LPM (heavier-than-air gases)
- Do not humidify span gas when flowing

Gas Detection Sensor Technologies

INFRARED

Toxic, combustible & refrigerant gas sensors

PID / TVOCS

OVERVIEW

Toxic gas sensors, organic compounds & chemicals

LIFE SPAN

10 years +

LIFE SPAN

3 - 8 years (typically if not contaminated and with regular maintenance)

SPECIFICITY

Non specific. Will respond to any compound that has an ionization potential less than the ionization potential of the lamp.

RANGE

0 - 1,000 ppm or 0 - 100% volume. Target gas, manufacturer dependent.

POISONING / DAMAGING

No known poisoning agents. Condensing humidity will damage sensor and distort readings.

APPLYING SPAN GAS

- Use span gas with air balance or nitrogen balance. Nitrogen balance ONLY for CO, sensors.
- Flow rate should be approximately 0.5 LPM. Some sensors are flow sensitive.
- Do NOT humidify span gas when flowing.

RANGE 0 -

0 - 30 ppm or 0 - 300 ppm, sensor dependent

CONTAMINATION

Many other gases, vapours, chemicals. Condensing humidity can cause false positive readings.

APPLYING SPAN GAS

- Use span gas with air or nitrogen balance.
- Flow rate should be a minimum of 0.5 LPM.
- Do NOT humidify span gas when flowing.

Specific to target gas.

SPECIFICITY

SPECIFICATIONS Target Gas Sensors

Releative Humidity (RH)

SENSOR		
Туре	Thin film capacitive	
Standard Range	5 - 95% RH (non-condensing)	
Resolution	2% RH	
Accuracy	No data available	
Long Term Drift	2% (±) / 12 months	
Response Time	< 10 seconds	

INSTRUMENT

Displayed Resolution	1%
Warm Up Time @ Switch On	5 minute operational, 20 minute max accuracy
Recommended Calibration Frequency	1 yr for best performance

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non- condensing
Operating Life (Estimated)	3 yrs +

Temperature

SENSOR		
Туре	Negative Coefficient Thermistor	
Standard Range	0°C to 50°C (32°F to 122°F)	
Resolution	0.1°C @ 25°C	
Accuracy	No data available	
Long Term Drift	0.5°C (±) / 12 months	
Response Time	< 10 seconds	

INSTRUMENT

Displayed Resolution	0.1°C
Warm Up Time @ Switch On	5 minute operational, 10 minute max accuracy
Recommended Calibration Frequency	1 yr for best performance

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non- condensing
Operating Life (Estimated)	6 yrs +

Ammonia (NH ₃)	50 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 50 ppm
Resolution	1 ppm
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{s_0} = < 60$ sec calculated fr 5 minute exposure
	Alcohols @ 1,000 ppm = 0 ppm
	CO ₂ @ 5,000 ppm = 0 ppm
	CO @100 ppm = 0 ppm
Cross Sensitivities	Hydrocarbons @ % range = 0 ppm
Closs Selisitivities	H ₂ @ 10,000 ppm = 0 ppm
	H,S @ 20 ppm = 2 ppm
	Cross sensitivity list not fully completed.
	Sensor maybe sensitive to other gases.

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non-condensing
Operating Life (Estimated)	2 yrs

Arsine (AsH ₃)	1 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 1 ppm
Resolution	< 15 ppb @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{q_0} = <30$ seconds calc fr 2 minute exposure
	CO @ 85 ppm = 0 ppm
	H ₂ @ 3,100 ppm = 0 ppm
Cross Sensitivities	$NO_2 @ 10 \text{ ppm} = 2 \text{ ppm}$
	$C_{3}H_{5}OH @ 25,000 \text{ ppm} = 0 \text{ ppm}$
	$H_2S @ 18 \text{ ppm} = 10.8 \text{ ppm}$

	SO ₂ @ 18 ppm = 5.4 ppm
	Cl, @ 0.85 ppm = 0.24 ppm
	HCI @ 7.8 ppm = 1 ppm
	HF @ 7.2 ppm = 0 ppm
Cross Sensitivities continued	HCN @ 12.6 ppm = 0.7 ppm
	SiH_{4} @ 4.3 ppm = 0.7 ppm
	$H_{2}Se @ 0.8 ppm = 0.24 ppm$
	$B_{2}H_{6}@0.2 \text{ ppm} = 0.28 \text{ ppm}$
	PH, @ 0.2 ppm = 0.24 ppm

Displayed Resolution	0.001 ppm (1 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	20 - 95% non- condensing
Operating Life (Estimated)	1.5 yrs

Carbon Dioxide (CO ₂)	5,000 ppm
SENSOR	
Туре	Infrared
Standard Range	0 - 5,000 ppm
Resolution	50 ppm fr 0 - 2,500 ppm, then 100 ppm up to FSD
Accuracy	\pm 2% full scale @ 20°C (68°F), 1 bar pressure, applied gas
	2.5% volume CO ₂
Long Term Drift	\pm 50 ppm / month @ 20°C (68°F) ambient, (max \pm 150 ppm
	/ yr)
Response Time	$t_{90} = > 30$ seconds @ 20°C (68°F)
Cross Sensitivities	None

INSTRUMENT

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condensing
Operating Life (Estimated)	5 - 10 yrs

Carbon Dioxide (CO ₂)	10,000 ppm
SENSOR	
Туре	Infrared
Standard Range	0 - 10,000 ppm
Resolution	50 ppm fr 0 - 2,500 ppm, then 100 ppm up to FSD
Accuracy	\pm 2% full scale @ 20°C (68°F), 1 bar pressure, applied gas 2.5% volume CO,
Long Term Drift	\pm 500 ppm / month @ 20°C (68°F) ambient
Response Time	$t_{q_0} = > 30$ seconds @ 20°C (68°F)
Cross Sensitivities	None

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condensing

Carbon Dioxide (CO ₂)	5% volume
SENSOR	
Туре	Infrared
Standard Range	0 - 5% volume
Resolution	1% of measuring range for readings above 50% of range, 0.5% of measuring range for readings below 50% of range
Accuracy	\pm 2% full scale @ 20°C (68°F), 1 bar pressure, applied gas 2.5% volume CO,
Long Term Drift	± 500 ppm / month @ 20°C (68°F) ambient
Response Time	$t_{q_0} = > 30$ seconds @ 20°C (68°F)
Cross Sensitivities	None
INSTRUMENT	
Displayed Pecalution	0.10/ volume

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condensing
Operating Life (Estimated)	5 - 10 yrs

Carbon Dioxide (CO ₂)	20% volume
SENSOR	
Туре	Infrared
Standard Range	0 - 20% volume
Resolution	0.1% volume
Accuracy	\pm 1% volume at STP(20°C, 101.325 kPa) & time of calibration
Long Term Drift	\pm 1% volume / month @ 20°C (68°F) ambient
Response Time	$t_{_{90}} = > 30$ seconds @ 20°C (68°F)
Cross Sensitivities	None

INSTRUMENT

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approx 2 - 2.5 min (warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condensing
Operating Life (Estimated)	5 - 10 yrs

Carbon Dioxide (CO ₂)	100% volume
SENSOR	
Туре	Infrared
Standard Range	0 - 100% volume
Resolution	1% of measuring range for readings above 50% of range,
Resolution	0.5% of measuring range for readings below 50% of range
	\pm 1% volume at STP (20°C, 101.325 kPa) & time of calibration
Accuracy	\pm 10% volume across temperature and pressure when
	calibrated at altitude
Long Term Zero Drift	\pm 1% volume / month @ 20°C (68°F) ambient
Response Time	$t_{q_0} = < 30$ seconds @ 20°C (68°F) ambient
Cross Sensitivities	None

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approx 3 min (warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	-20°C to 50°C (-4°F to 122°F)	
Operating Humidity	0 - 95% non-condensing	
Operating Life (Estimated)	> 5 yrs	

Carbon Monoxide (CO)	50 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 50 ppm
Resolution	0.5 ppm
Accuracy	No data available
Long Term Drift	zero: 0.2 ppm equivalent change / yr in clean air. Sensitivity: 3% change / yr in clean air (value based on twice per month test)
Response Time	$t_{g_0} = < 25$ seconds fr 0 - 400 ppm
Cross Sensitivities	H ₂ S @ 20 ppm = < 0.1 ppm NO ₂ @ 10 ppm = < 0.1 ppm Cl ₂ @ 10 ppm = < 0.1 ppm NO @ 50 ppm = < 5 ppm
Cross Sensitivities <i>continued</i>	$SO_{2} @ 20 \text{ ppm} = < 0.1 \text{ ppm}$ $H_{2} @ 20^{\circ}C (68^{\circ}F) @ 400 \text{ ppm} = < 60 \text{ ppm}$ $C_{2}H_{4} @ 400 \text{ ppm} = < 25 \text{ ppm}$ $NH_{3} @ 20 \text{ ppm} = < .01 \text{ ppm}$

INSTRUMENT

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	2 yrs, 1 yr for best accuracy

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non-condensing
Operating Life (Estimated)	2 - 3 yrs

Carbon Monoxide (CO)	50 ppm
SENSOR	
Туре	Electrochemical (for use in H ₂ background environment)
Standard Range	0 - 50 ppm
Resolution	0.5 ppm
Accuracy	No data available
	zero: 0.2 ppm equivalent change / yr in clean air.
Long Term Drift	Sensitivity: 3% change / yr in clean air (value based on twice
	per month test)
Response Time	$t_{90} = < 30$ seconds
	H_{2}^{2} @ 900 ppm in 900 ppm CO @ 10°C (50°F) = < 2 ppm
	$H_{2,0}^{2}$ @ 900 ppm in 900 ppm CO @ 20°C (68°F) = < 4 ppm
	H, @ 900 ppm in 900 ppm CO @ 30°C (86°F) = < 6 ppm
	NO ₂ , @ 10 ppm = < 0.1 ppm
Cross Sensitivities	Cl, @ 10 ppm = < 0.1 ppm
	NO @ 50 ppm = < 0.1 ppm
	SO ₂ @ 20 ppm = < 0.1 ppm
	$C_{2}H_{4} @ 400 \text{ ppm} = < 30 \text{ ppm}$
	NH ₃ @ 20 ppm = < 0.1 ppm

Displayed Resolution	0.5 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non-condensing
Operating Life (Estimated)	2 - 3 yrs

Chlorine (Cl ₂)	5 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 5 ppm
Resolution	0.02 ppm
Accuracy	No data available
Long Term Drift	zero: < 0.2 ppm equivalent change / yr in clean air with monthly test. Sensitivity: < 0.4 ppm change / month in clean air with twice monthly test

Response Time	$t_{so} = < 40$ seconds fr 0 - 5ppm (diffusion)
	H,S @ 20 ppm= < -40 ppm
	NO, @ 10ppm= 100 ppm
	NO @ 50 ppm = < 0.5 ppm
Cross Sensitivities	SO ₂ @ 20 ppm = < -2.5 ppm
	CO @ 400 ppm = < 0.1 ppm
	H ₂ @ 400 ppm = < 0.1 ppm
	$C_{2}H_{4}@400 \text{ ppm} = < 0.1 \text{ ppm}$

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non-condensing
Operating Life (Estimated)	2 - 2.5 yrs

Chlorine Dioxide (ClO ₂)	1 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 1 ppm
Resolution	0.02 ppm
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{s_0} = $ < 20 seconds calc fr 2 minute exposure time
Response fille	$t_{90} = < 120$ seconds calc fr 2 minute exposure
	Alcohols @ 1,000 ppm = 0 ppm
	CO @ 100 ppm = 0 ppm
Cross Sensitivities	Cl ₂ @1ppm = 0.6 ppm
Closs Selisitivities	0 ₃ @ 0.25 ppm = 0.7 ppm
	H_2^{-} @ 3,000 ppm = 0 ppm
	$H_{2}^{2}S @ 20 \text{ ppm} = -5 \text{ ppm}$

INSTRUMENT

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 95% non- condensing
Operating Life (Estimated)	2 yrs

Combustibles	100% LEL
SENSOR	
Туре	Catalytic Pellistor
Standard Range	0 - 100% LEL
Resolution	1% LEL
Accuracy	No data available
Long Term Drift	Minimal
Response Time	$t_{so} = < 10$ seconds
Cross Sensitivities	Responds to most flammable gases & vapours

INSTRUMENT

Displayed Resolution	1% LEL
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	1 yr for best performance

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	10 - 90% non- condensing
Operating Life (Estimated)	5 yrs +

Combustibles	5% volume CH_4
SENSOR	
Туре	Infrared
Standard Range	0 - 5% vol CH ₄
Resolution	0.1% vol CH_4
Accuracy	No data available
Long Term Drift	\pm 1% FSD / mth @ 20°C (68°F) ambient, (max \pm 3% of full
	scale / yr)
Response Time	$t_{ao} = <30$ seconds @ 20°C (68°F) ambient
Cross Sensitivities	None

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	1 year

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	5 - 95% non-condencing
Operating Life (Estimated)	5 - 10 yrs

Ethylene (C ₂ H ₄)	200 ppm	
SENSOR		
Туре	Electrochemical	
Standard Range	0 - 200 ppm	
Resolution	1 ppm	
Accuracy	No data available	
Long Term Drift	< 5% / month	
Response Time	$t_{q_0} = < 100$ seconds	
Cross Sensitivities	$column{0}{0} = < 60\%$	

INSTRUMENT

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs

Ethylene Oxide (C ₂ H ₄ O)	20 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 20 ppm
Resolution	0.1 ppm
Accuracy	No data available
Long Term Drift	< 5% signal loss / yr
Response Time	$t_{y_0} = < 120$ seconds

Cross Sensitivities	Ethanol \approx 55%
	Toluene \approx 20%
	Methyl-ethyl-ketone \approx 10%
	$C0 \approx 40\%$

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs

Fluorine (F ₂)	2 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 2 ppm
Resolution	< 0.02 ppm @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 5% / month
Response Time	$t_{q_0} = $ < 80 sec calc fr 4 minute exposure with 1 ppm Cl ₂
	Alcohols @ 1,000 ppm = 0 ppm
	AsH ₃ @ 0.2 ppm = -0.03 ppm
	Br = yes; n/d
	CO ₂ @ 5,000 ppm = 0 ppm
	CO @ 100 ppm = 0 ppm
	Cl ₂ @ 1 ppm = 1.4 ppm
	$B_2 H_6 @ 0.25 \text{ ppm} = -0.01 \text{ ppm}$
	Hydrocarbons @ % range = 0 ppm
Cross Sensitivities	HCl @ 5 ppm = -7 ppm
	H ₂ @ 10,000 ppm = 0 ppm
	HCN @ 1 ppm = -0.05 ppm
	H,S @ 1 ppm = -2 ppm
	$N_{2}^{2} @ 100\% = 0 \text{ ppm}$
	NO, @ 10 ppm = 8 ppm
	$0_{3} @ 0.25 \text{ ppm} = 0.3 \text{ ppm}$
	$PH_{2} @ 0.3 ppm = approximately -0.1 ppm; n/d$
	S0, @ 20 ppm = -0.2 ppm

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	1.5 - 2 yrs

Formaldehyde (CH ₂ O)	10 ppm	See Important Notes # 4
SENSOR		
Туре	Electrochemical	
Standard Range	0 - 10 ppm	
Resolution	0.01 ppm	
Accuracy	No data available	
Long Term Drift	< 2% signal loss / month	
Response Time	t _{so} = < 80 sec	
	$H_2 = 1 - 3\%$	
Cross Sensitivities	CO = 10 - 18%	
	Interference from other reducing	gases such as alcohol.

INSTRUMENT

Displayed Resolution	0.01ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs in air

Hydrogen (H,)	2,000 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 2,000 ppm
Resolution	2 ppm
Accuracy	No data available
Long Term Drift	< 2% / months
Response Time	$t_{q_0} = < 90$ seconds
Cross Sensitivities	$CO @ 300 \text{ ppm} = \le 60 \text{ ppm}$
	H ₂ S @ 15 ppm = < 3 ppm
	$SO_2 @ 5 ppm = 0 ppm$
	NO @ 35 ppm \approx 10 ppm
	$NO_2 @ 5 ppm = 0 ppm$
	$Cl_2 @ 1 ppm = 0 ppm$
	$HCN @ 10 \text{ ppm} \approx 3 \text{ ppm}$
	HCl @ 5 ppm = 0 ppm
	C_2H_4 @ 100 ppm \approx 80 ppm

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 yrs +

Hydrogen Chloride (HCl)	30 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 30 ppm
Resolution	< 0.7 ppm @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 3% / month
Response Time	$t_{q_0} = < 70$ seconds calc fr 4 minute exposure
Cross Sensitivities	Alcohols @ 1,000 ppm = 0 ppm
	NH ₃ @ 100 ppm = 0 ppm
	$AsH_{3} @ 0.2 \text{ ppm} = 0.7 \text{ ppm}$
	$CO_2 = 5,000 \text{ ppm} = 0 \text{ ppm}$

C0 @ 100 ppm = 0 ppm
$Cl_2 @ 5 ppm = < \pm 0.1 ppm$
Hydrocarbons @ % range = 0 ppm
H ₂ @ 10,000 ppm = 0 ppm
HČN @ 20 ppm = 7 ppm
H ₂ S @ 20 ppm = 60 ppm
NO @ 100 ppm = 45 ppm
N ₂ @ 100% = 0 ppm
$N_{0_{2}}^{0}$ @ 10 ppm = < ±0.5 ppm
PH ₃ @ 0.1 = 0.3 ppm
$SO_{2}^{2}@20 \text{ ppm} = 8 \text{ ppm}$

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 95% non- condensing
Operating Life (Estimated)	2 yrs

Hydrogen Cyanide (HCN)	30 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 30 ppm
Resolution	0.2 ppm
Accuracy	No data available
Long Term Drift	< 5% / month
Response Time	$t_{q_0} = < 50$ seconds calc fr 2 minute exposure
Cross Sensitivities	Alcohols @ 1,000 ppm = 0 ppm $CO_2 @ 5,000 ppm = 0 ppm$ CO @ 100 ppm = 0 ppm Hydrocarbons @ % range = 0 ppm $H_2 @ 10,000 ppm = 0 ppm$ NO @ 100 ppm = -5 ppm $NO_2 @ 10 ppm = -7 ppm$ $H_2 S @ 20 ppm = 0 ppm$ (short gas exposure in minute range; after filter saturation: ca. 40 ppm reading)

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 95% non- condensing
Operating Life (Estimated)	1.5 - 2 yrs

Hydrogen Fluoride (HF)	10 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 10 ppm
Resolution	0.2 ppm @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 10% in 6 months
Response Time	$t_{q_0} = < 90$ seconds calc fr 4 minute exposure
	$C_{2}H_{4}O_{2}@$ 100 ppm = 100 ppm
	Alcohols @ 1,000 ppm = 0 ppm
	CO ₂ @ 5,000 ppm = 0 ppm
	CO @ 100 ppm = 0 ppm
Cross Sensitivities	Cl ₂ @ 1 ppm = 0.7 ppm
	Hydrocarbons @ % range = 0 ppm
	H ₂ @ 3,000 ppm = < 1 ppm
	HCl @ 10 ppm = 6 ppm
	SO ₂ @ 20 ppm = 16 ppm

INSTRUMENT

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	1.5 yrs +

Hydrogen Sulphide (H ₂ S)	50 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 50 ppm
Resolution	< 0.05 ppm
Accuracy	No data available
Long Term Drift	Zero: < 0.1 ppm equivalent change / yr in clean air
	Sensitivity: < 4% change / yr in clean air with monthly test
Response Time	t ₉₀ = < 25 seconds fr 0 - 20 ppm
	$NO_{2} @ 10 \text{ ppm} = < -20 \text{ ppm}$
	Cl, @ 10 ppm = -25 ppm
	NÕ @ 50 ppm = < 4 ppm
Cross Sensitivities	SO ₂ @ 20 ppm = < 10 ppm
	CO @ 400 ppm = < 1.5 ppm
	H ₂ @ 400 ppm = < 0.2 ppm
	$C_{2}H_{4}@400 \text{ ppm} = < 0.5 \text{ppm}$
	$NH_{3} @ 20 \text{ ppm} = < 0.1 \text{ ppm}$

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)	
Operating Humidity	15 - 90% non- condensing	
Operating Life (Estimated)	2 - 3 yrs	

Nitrogen Dioxide (NO ₂)	10 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 10 ppm
Resolution	0.1 ppm
Accuracy	No data available
Long Term Drift	< 2% signal loss / month
Response Time	$t_{q_0} = < 25$ seconds
	H ₂ S @ 20 ppm = < -40 ppm
	$(\bar{U_{2}} @ 10 \text{ ppm} = 100 \text{ ppm})$
	NO @ 50 ppm = < 0.5 ppm
Cross Sensitivities	SO ₂ @ 20 ppm = < -2.5 ppm

	CO @ 400 ppm = < 0.1 ppm
	H ₂ @ 400 ppm = < 0.1 ppm
Cross Sensitivities continued	$C_{2}H_{4}$ @ 50 ppm = < 0.1 ppm
	NH, @ 20 ppm = < 0.1 ppm
	$CO_{2} @ 5\%$ volume = < 0.1 ppm

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs

Nitric Oxide (NO)	100 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 100 ppm
Resolution	< 0.2 ppm
Accuracy	No data available
Long Term Drift	Zero: 0.5 ppm equivalent change fr -20°C to 20°C (-4°F to 68°F), 1 - 3 ppm equivalent change 20°C to 50°C (68°F to 122°F) Sensitivity: 101 - 105% output change @ 50 ppm btw 20°C (68°F) & 50°C (122°F)
Response Time	$t_{_{90}} = $ < 20 seconds fr 0 - 50 ppm
Cross Sensitivities	$H_2S @ 20 ppm = < 30 ppm$ $NO_2 @ 50 ppm = < 5 ppm$ $CI_2 @ 10 ppm = < 15 ppm$ $SO_2 @ 20 ppm = < 3 ppm$ $H_2 @ 400 ppm = < 0.1 ppm$ CO @ 400 ppm = < 0.1 ppm $NH_3 @ 20 ppm = < 0.1 ppm$ $CO_2 @ 5\% volume = < 0.1 ppm$

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

ENVIRONMENTAL

Operating Temperature	0°C to 50°C (32°F to 122°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 - 3 yrs

Oxygen (O ₂)	25% volume
SENSOR	
Туре	Electrochemical
Standard Range	0 - 25% volume
Resolution	0.1% volume
Accuracy	No data available
Long Term Drift	< 1% change in output over 3 months
Response Time	$t_{q_0} = < 15$ seconds fr 0 - 20.9%
Cross Sensitivities	\widetilde{CO}_2 sensitivity: 0.1% change in O_2 reading per % CO ₂ in 5% CO ₂

INSTRUMENT

Displayed Resolution	0.1% volume
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 55°C (32°F to 131°F)
Operating Humidity	< 0.7% fr 0 - 95% RH @ 40°C (104°F)
Operating Life (Estimated)	3 yrs

Ozone (0 ₃)	1 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 1 ppm
Resolution	< 0.02 ppm @ 20°C (68°F)
Accuracy	No data available
Long Term Drift	< 10% / 6 months @ 20°C (68°F) and 30 - 50% RH
	$t_{q_0} = < 60$ seconds calc from 3 minute exposure
Response Time	@ 30 cc / min flow
	Br, $I_2 = yes; n/d$
	$CO_2 = 5,000 \text{ ppm} = 0 \text{ ppm}$
	CO @ 100 ppm = 0 ppm
Cross Sensitivities	Cl ₂ @1ppm = 1.2 ppm
	$N_2H_4@3 ppm = -3 ppm$
	H ₂ @ 3,000 ppm = 0 ppm
	$H_2^{-}S @ 20 \text{ ppm} = -1.6 \text{ ppm}$
	N ₂ @ 100% = 0 ppm
	NO ₂ @ 10ppm = 6 ppm

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non-condensing
Operating Life (Estimated)	1.5 - 2 yrs

Phosphine (PH ₃)	5 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 5 ppm
Resolution	Lower detection limit < 30 ppb
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{90} = < 30$ seconds
	CO @ 85 ppm = 0 ppm
	H ₂ @ 3,100 ppm = 0 ppm
	NO ₂ @ 10 ppm = 2 ppm
	C ₃ H ₅ OH @ 25,000 ppm = 0 ppm
	H ₂ S @ 18 ppm = 13 ppm
	SO ₂ @ 18 ppm = 6.5 ppm
Cross Sensitivities	Cl ₂ @ 0.85 ppm = 0.29 ppm
Closs Selisitivities	HČl @ 7.8 ppm = 1.2 ppm
	HF @ 7.2 ppm = 0 ppm
	HCN @ 12.6 ppm = 0.84 ppm
	SiH ₄ @ 4.3 ppm = 0.84 ppm
	H,Se @ 0.8 ppm = 0.29 ppm
	$B_{2}H_{6}@0.2 \text{ ppm} = 0.34 \text{ ppm}$
	$A_{SH_{3}}^{*}$ @ 0.2 ppm = 0.16 ppm

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

0°C to 40°C (32°F to 104°F)
10 - 95%
non- condensing
2 yrs

Silane (SiH ₄)	20 ppm
SENSOR	
Туре	Electrochemical
Standard Range	0 - 20 ppm
Resolution	0.05 ppm
Accuracy	No data available
Long Term Drift	< 5% / 6 months
Response Time	$t_{ao} = < 60$ seconds calc fr 2 min exposure
Cross Sensitivities	CO @ 85 ppm = 0 ppm H ₂ @ 3,100 ppm = 0 ppm NO ₂ @ 10 ppm = 2.3 ppm C ₃ H ₅ OH @ 25,000 ppm = 0 ppm H ₂ S @ 18 ppm = 8 ppm SO ₂ @ 18 ppm = 7.4 ppm Cl ₂ @ 0.85 ppm = 0.1 ppm HCl @ 8 ppm = 0.45 ppm HF @ 7.2 ppm = 0 ppm HCN @ 12 ppm = 0.77 ppm AsH ₃ @ 0.16 ppm = 0.2 ppm H ₂ Se @ 0.8 ppm = 0.2 ppm B ₂ H ₆ @ 0.2 ppm = 0.27 ppm PH ₃ @ 0.2 ppm = 0.35 ppm

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	20 - 95% non- condensing
Operating Life (Estimated)	1.5 yrs

Sulphur Dioxide (SO ₂)	20 ppm	
SENSOR		
Туре	Electrochemical	
Standard Range	0 - 20 ppm	
Resolution	< 0.1 ppm	
Accuracy	No data available	
Long Term Drift	< 2% change / month in clean air	
Response Time	$t_{g_0} = < 25$ seconds fr 0 - 10 ppm	
	$H_{2}S @ 20 \text{ ppm} = < 0.1$	
	$NO_{2}@$ 10 ppm = < -130 ppm	
	Cl ₂ @ 10 ppm = < -40 ppm	
Cross Sensitivities	NO @ 50 ppm = < ±2 ppm	
Closs Selisitivities	CO @ 400 ppm = < 1.6 ppm	
	H ₂ @ 400 ppm = < 0.3 ppm	
	$C_{2}H_{4} @ 400 \text{ ppm} = < 40 \text{ ppm}$	
	NH ₃ @ 20 ppm = < 0.1 ppm	

Displayed Resolution	0.1 ppm (100 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	6 months

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	15 - 90% non- condensing
Operating Life (Estimated)	2 yrs

Total Volatile Organic Compound (TVOC)

30 ppm

SENSOR

Туре	Photolonization Detector
Standard Range	0 - 30 ppm
Resolution	0.02 ppm
Accuracy	No data available
Long Term Drift	< 2% change / month in clean air
Response Time	$t_{ao} = < 3$ seconds
Cross Sensitivities	Many chemicals & gases. Refer to manual.

INSTRUMENT

Displayed Resolution	0.01 ppm (10 ppb)
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	monthly to 4 months (usage dependent)

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	5 - 95% non- condensing
Operating Life (Estimated)	5 yrs (excluding replacable lamp & electrode stack)

Total Volatile Organic Compound (TVOC)		300 ppm	
SENSOR			
Туре	Photolonization Detec	tor	
Standard Range	0 - 300 ppm		
Resolution	0.1 ppm		
	(100 ppb)		
Accuracy	No data available		
Long Term Drift	< 2% change / month in clean air		
Response Time	$t_{q_0} = <3$ seconds		
Cross Sensitivities	Many chemicals & gas	es. Refer to manual.	
INSTRUMENT			
Dicplayed Pecalution	1 nnm		

Displayed Resolution	1 ppm
Warm Up Time @ Switch On	Approximately 2 - 2.5 min (instrument warm up delay time)
Recommended Calibration Frequency	monthly to 4 months (usage dependent)

ENVIRONMENTAL

Operating Temperature	0°C to 40°C (32°F to 104°F)
Operating Humidity	5 - 95% non- condensing
Operating Life (Estimated)	5 yrs (excluding replacable lamp & electrode stack)

Additional chemical symbols not defined above:

Br	Bromine	C,H,	Acetylene
CIF,	Chlorine Trifluoride	C, H, 0	Isopropyl Alcohol

B,H,

Important Notes:

- Some sensors may be calibrated with correlation gases. If you prefer to have specific sensors
 calibrated with the target gas, contact our factory for availability and extra costs. Customer will have
 to bear the cost of the full cylinder of specialty gas plus incoming dangerous goods freight and take
 ownership of the cylinder of gas remaining.
- 2. These specifications have been developed from data considered accurate at the time. No warranty is implied or suggested based on this data. We accept no responsibility for errors or omissions.
- 3. Critical Environment Technologies Canada Inc. reserves the right to make design and specification changes without prior notice.
- 4. Formaldehyde sensor has high cross sensitivity to Carbon Monoxide, Alcohol & Hydrogen.
- 5. Combustible (flammable) gas sensors (catalytic) can be calibrated for a number of target gases. Please specify the target gas desired & we will evaluate your request.
- 6. Chlorine, Fluorine, Hydrogen Chloride, Hydrogen Fluoride and Ozone gas sensors are not available with the YES AIR Pump model.



Unit 145, 7391 Vantage Way, Delta, BC, V4G 1M3, Canada

Toll Free: +1.877.940.8741 Tel: +1.604.940.8741 Fax: +1.604.940.8745

www.critical-environment.com





© April 2015 Critical Environment Technologies Canada Inc. All rights reserved. Data in this publication may change without notice.